VeriGUIDE™ 3D Print Resin
Instruction for Use (IFU) – For Professional Use Only

Indications for Use

Whip Mix VeriGUIDE™ is a 3D print resin for producing drilling guides used in dental surgery. VeriGUIDE™ is compatible with DLP and LCD printers featuring 385nm and 405nm light sources.

Storage and Transportation

1. VeriGUIDE™ must be stored in the original packaging at room temperature in a dry area away from sunlight.
2. Always keep container tightly sealed and close the container immediately after each use.
3. Do not expose VeriGUIDE™ to any light source. A minimal amount of light can induce polymerization (solidification).
4. Resin that is stored in a tray must be stirred before each print using a plastic or paper card (similar to a business card), to ensure a homogenous mixture.
5. Do not use the product past the expiration date.
6. The lot number and expiration date are indicated on VeriGUIDE™ packaging. In case of questions or concerns, please refer to the lot number.
7. Packages are ideally transported at room temperature and away from light in a tightly sealed container to prevent leakage.

Design Recommendation

Wall Thickness	**2 mm**
Internal offset	**≥ 0.1 mm**
Block out angle	**0°**
Retention amount	**0 mm**
Offset from Sleeve	**0.10 mm**

Validated Equipment

For all validated printers, curing units, and their setting, please scan the QR code

CAUTION: Only use the validated equipment (3D printers and post-curing units) specified on the Whip Mix website. Substitution or modifications may result in a device that does not meet safety and performance standards and could pose risk to patients.

Processing

1. When printing with VeriGUIDE™ 3D Print Resin, make sure the resin is at a temperature between 20°C and 30°C. Extreme high or low temperatures could potentially affect the accuracy of the printed models or cause failure of the printed objects.
2. Wear proper PPEs when handling uncured resin.
3. VeriGUIDE™ bottle must be shaken thoroughly before use.
4. Don't add supports to the intaglio surface.

5. Don’t print with the end of the surgical guide arch attached to the build plate with any print orientation because failure could happen at the tip of the arch.
6. After the build process is finished, post processing is required.

Post Processing

1. When the print job is complete, remove the build platform from the printer, and then carefully remove the printed objects from the platform.
2. The printed object must be rinsed with Isopropyl alcohol using a washing unit for 5 minutes to eliminate any excess resin.
3. Printed objects must be rinsed with Isopropyl alcohol in a second fresh washing unit for an additional 5 minutes.

NOTE: Do not immerse the printed parts in alcohol for more than 10 minutes, as prolonged exposure may cause the material to deteriorate.

4. To remove excess alcohol, leave prints to dry at room temperature (or compressed air can also be used).
5. Place the printed objects in the light curing unit such as Otoflash G171 for 6,000 flashes (scan QR code for recommended curing units).
6. Look for cured debris in the tray after each print job. A paint strainer can be used to filter out any debris. Printing with cured debris could potentially cause damage to the tray or the printed objects.
7. Use crosscut carbide bur at a slow speed to remove the remaining printed support pins.
8. Deviations from the described manufacturing process may lead to different mechanical and optical properties of the material.
9. The surgical guide can be disinfected and is compatible with sterilization methods such as steam/autoclave (up to 135°C for 5 minutes), ethylene oxide, and gamma irradiation. Be sure to follow sterilization guidelines. Please note that gamma irradiation may cause discoloration.

Warning!

(Covering US, Mexico, Canada and Australia)

Contains Methacrylate oligomers and monomers, Acrylate oligomers and monomers, and Phosphine Oxide.

H315 Causes skin irritation.
H317 May cause an allergic skin reaction.
H319 Causes serious eye irritation.
H335 May cause respiratory irritation.
H360 May damage fertility or the unborn child.
H411 Toxic to aquatic life with long lasting effects.

Precautionary Phrases

P201 Obtain special instructions before use.
P202 Do not handle until all safety precautions have been read and understood.
P261 Avoid breathing mists, vapors or spray.
P264 Wash thoroughly after handling.
P271 Use only outdoors or in a well-ventilated area.
P272 Contaminated work clothing must not be allowed out of the workplace.
P273 Avoid release to the environment.
P280 Wear protective gloves and eye protection.
P308 + P313 IF exposed or concerned: Get medical attention.
P302 + P352 IF ON SKIN: Wash with plenty of soap and water.
P333 + P313 If skin irritation or rash occurs: Get medical attention.
P362 + P364 Take off contaminated clothing and wash it before reuse.
P304 + P340 IF INHALED: Remove person to fresh air and keep comfortable for breathing.
P312 Call a POISON CENTER or doctor if you feel unwell.

P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.
P337 + P313 If eye irritation persists: Get medical attention.
P391 Collect spillage.
P403 + P233 Store in a well-ventilated place. Keep container tightly closed.
P405 Store locked up.
P501 Dispose of contents and container in accordance with local and national regulations.

Symbols and their Description

Symbol	Description
	See Instructions for use
	Manufacture by
	Use-by-date
	Keep away from sunlight
	Temperature limit
	Shake well before use
	Batch number
	Unique Device Identifier
	Reference number
	Prescription-only

Whip Mix Corporation, 361 Farmington Avenue, Louisville, Kentucky, USA, 40209

Phone: 502- 637-1451, www.whipmix.com

SRL0020 Rev C 6/2026